## **<u>"TKA-KEEPER"</u>**



## Main technical characteristics

| IIIIIMINATION MARCURAMANT RANGA                    | 10 ÷<br>200,000 lx     |
|--|------------------------|
| Basic relative error of illumination measurements  | ± 8.0%                 |
| <b>EVALUATE TRADES OF TRADIANCE</b> $IV = (A + B)$ | 1 ÷ 200,000<br>mW / m² |
| IMPASULEMENT FANDE OF IFEADIANCE IV. (A + B)       | 10 ÷ 40,000<br>mW / m² |
| Basic relative error of irradiance measurements    | ± 10.0%                |
| Air temperature measurement range                  | -30 ÷ +60 °<br>C       |

|   | 1  |
|---|--|
| Basic absolute error of temperature measurements  | ± 0.2 ° C  |
| Limits of additional absolute error of air temperature<br>measurements at air temperature, ° C<br>from -30 to -10 inclusive<br>above -10 to +15 inclusive<br>above +25 to +45 inclusive<br>above +45 to +60   | ± 0.3 ° C<br>± 0.1 ° C<br>± 0.1 ° C<br>± 0.3 ° C |
| Measurement range of <b>relative air humidity</b>   | 5 ÷ 98%  |
| Basic absolute error of relative humidity measurements  | ± 3.0% rel.<br>ow.                               |
| Limits of the additional absolute error of relative humidity<br>measurements when the air temperature changes by every 10 ° C in<br>the range from +10 to +60 ° C   | ± 3.0% rel.<br>ow.                               |
| The limits of the additional relative error of the device when<br>measuring optical quantities, due to a change in the sensitivity<br>of the photometric head when the air temperature in the<br>measurement zone changes for every 10 °C (no more) | ± 3.0%   |

## Dimensions

| <ul> <li>signal processing unit (no more)</li> </ul>        | 130  | Х  | 70 | Х | 30 | mm |
|---|------|----|----|---|----|----|
| <ul> <li>photometric head with a probe (no more)</li> </ul> | 230  | Х  | 48 | Х | 55 | mm |
| Device weight (no more)                                     | 0,4  | kç | ļ  |   |    |    |
| Battery — Krona battery standard size                       | 9 in | า  |    |   |    |    |

The difference in the function of the relative spectral sensitivity of the photodetector of the Luxmeter is corrected by a system of light filters to match the function of the relative spectral luminous efficiency of monochromatic radiation for daytime vision V ( $\lambda$ ) according to GOST 8.332.



The effective reference plane of the Luxmeter coincides with the front plane of the cosine attachment of the photodetector

## Substantial benefits

It combines the functions of a universal light meter, a UV radiometer with increased sensitivity, and a temperature and humidity meter. Faster response, improved performance, high precision platinum temperature sensor, low power consumption.

"TKA-KHANITEL" is a multifunctional measuring complex for research and

operational control of microclimate parameters affecting objects of sociocultural and artistic-historical significance in museums, archives, warehouses, for use in cultural and art institutions, school and local history clubs.

- Operation manual "TKA-KEEPER"
- Factory calibration in accordance with MP-242-1969-2016, approved by the State Research Center for SI "VNIIM im. DI. Mendeleev "November 26, 2016